



Ref. Certif. No.

DE 3 - 30624

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE**
CERTIFICAT D'ESSAI OC

Product

Measuring appliances
(9 Channels Power Analyzer)

Name and address of the applicant

ENERGY TEAM S.P.A.
Via della Repubblica 9
20090 Trezzano sul Naviglio (MI), ITALY

Name and address of the manufacturer

ENERGY TEAM S.P.A.
Via della Repubblica 9, 20090 Trezzano sul Naviglio (MI), ITALY

Name and address of the factory

ENERGY TEAM S.P.A.
Via della Repubblica 9, 20090 Trezzano sul Naviglio (MI), ITALY

Ratings and principal characteristics

Rated Voltage:	24-120 Vdc, 100-240 VAC
Rated Power:	2 W
Rated Frequency:	50-60 Hz
Protection Class:	II
Degree of protection against liquids:	IP20
Measurement category:	III 240 VAC

Trade mark (if any)

Energy Team

Model/type Ref.

NG-9, NG-9 PLUS

Additional information (if necessary)

A sample of the product was tested and found
to be in conformity withIEC 61010-1:2010
IEC 61010-2-030:2010as shown in the Test Report Ref. No.
which forms part of this certificate

068-ELS1008714A-000

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**Date, 2018-01-17
CB 17 12 03296 002

Domenico Pignataro



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service