

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB EMC TEST CERTIFICATE

Product

Measuring appliances
(9 Channels Power Analyzer)

Name and address of the applicant

ENERGY TEAM S.P.A.
Via della Repubblica 9
20090 Trezzano sul Naviglio (MI), ITALY

Name and address of the manufacturer

ENERGY TEAM S.P.A.
Via della Repubblica 9, 20090 Trezzano sul Naviglio (MI), ITALY

Name and address of the factory

ENERGY TEAM S.P.A.
Via della Repubblica 9, 20090 Trezzano sul Naviglio (MI), ITALY

Ratings and principal characteristics

Rated Voltage:	24-120 Vdc, 100-240 VAC
Rated Power:	2 W
Rated Frequency:	50-60 Hz
Protection Class:	II
Degree of protection against liquids:	IP20
Measurement category:	III 240 VAC

Trade mark (if any)

Energy Team

Model/type Ref.

NG-9, NG-9 PLUS

Additional information (if necessary)

A sample of the product was tested and found to be in conformity with

IEC 61326-1:2012

as shown in the Test Report Ref. No.
which forms part of this certificate


TÜV SÜD Product Service
068-EMC1008714A-000

This CB EMC Test Certificate is issued by the National Certification Body

Date,

2018-01-17

CBE 17 12 03296 001



(Domenico Pignataro)



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service